



Voyant Maryland Fusion Device with Single-Step Activation

DEVICE GUIDE


Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

OVERVIEW

The Voyant Maryland Fusion device with single-step activation is a bipolar, electro-surgical device used to effectively seal and divide vessels up to and including 7mm in diameter and tissue bundles that can be captured in the jaws of the device. The device features a spring-loaded handle with an inset fuse activation button that reduces the steps required to deliver energy.

Unrestricted Shaft Rotation

Allows for continuous, 360-degree rotation to facilitate accurate jaw placement.



Single-Action, Curved, Tapered Jaw with Dissecting Tip

Aids in tracking the contour of anatomical structures, visualizing the tips of the jaw, and optimizing control during dissection.

Produces enhanced tissue dissection.

Blade Lever

Transects tissue with a mechanical blade.



Spring-Loaded Handle Lever with Inset Fuse Activation Button

Increases user efficiency by reducing the steps required for energy activation.


Opens automatically when released.

Device Specifications


Model	Shaft Length	Trocar Compatibility	Seal Length	Cut Length
EB212	37cm	5mm or larger	20mm	18mm
EB213	44cm	5mm or larger	20mm	18mm
EB214	23cm	5mm or larger	20mm	18mm

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION DEMONSTRATION

Note: This guide is not intended to replace the Instructions for Use (IFU). Please refer to the IFU for the indications, contraindications, warnings, precautions, instructions and other information.



- 1 Position the jaw using the 360-degree rotation knob.




- 2 Grasp the tissue or vessel by compressing the handle lever through the grasping zone. When the lever reaches the end of the grasping zone, the jaws will have closed.

WARNINGS:

- Do not place the tissue or vessel in the jaw hinge. Place it in the center of the jaw.
- If intending only to grasp or manipulate tissue or vessels, do not advance the handle lever past the grasping zone, as this will deliver energy.

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

DEMONSTRATION




- 3** When you are ready to activate energy, compress the handle lever past the grasping zone to depress the fuse activation button.

WARNINGS*:

- Do not attempt to seal vessels greater than 7mm in diameter with this device.
- During device activation, seal 1–3mm away from unintended thermally sensitive structures. To prevent unintended thermal injury, care should be taken around structures in close proximity to the jaws and during surgical procedures occurring in confined spaces.
- Keep handle lever fully closed throughout the entire seal cycle to ensure complete sealing. The handle lever does not latch and must be held closed during the duration of the seal cycle.

* Please refer to page 10 for additional warnings.

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION DEMONSTRATION



- 4** Transect tissue by using your index finger to pull the blade lever.

WARNINGS:

- Do not deploy the blade until the seal cycle is complete.
- Maintain pressure on the handle lever while transecting tissue or a vessel, to avoid accidental reactivation of energy.




- 5** Release the blade and handle levers to open the jaws.

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

ADDITIONAL IN-SERVICE RECOMMENDATIONS


Activating Only Once When Sealing

The Voyant algorithm recognizes the tissue between the jaws. The device needs to be activated only once to create a permanent, fused seal. Multiple activations on the same seal site are not necessary and not recommended per the IFU.




WARNING:


Do not seal directly over an area that is already sealed, or the integrity of the primary seal may be compromised.




Place the tissue or vessel in the center of the jaw, not in the jaw hinge. Placing the tissue in the hinge can lead to an incomplete seal.



When grasping the tissue or vessel, ensure that the angle of the jaw is **perpendicular** to the vessel or tissue. A seal made at an **acute angle** cannot withstand as much systolic pressure.



Jaws overfilled at the hinge




Potential for an incomplete cut

WARNING:


Do not overfill the jaw of the device with tissue. Doing so could compromise the device's sealing and cutting functions, prevent the jaw from opening properly, or result in an incomplete cut.

Sealing Without Applying Tension

To ensure proper function of the device, do not apply tension while sealing and transecting tissue.



No tension (recommended)



High tension (not recommended)

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

ADDITIONAL IN-SERVICE RECOMMENDATIONS

Cleaning the Jaws


- For optimal performance, keep the blade channel and jaw surfaces clean. With the jaws closed, pull and release the blade lever to clear the blade channel of eschar buildup.
- Use a gauze pad soaked with sterile water or saline to remove eschar.



1 Release the handle lever to open the jaws.



2 Wet a gauze pad with saline.



3 Wipe the open jaws clean with the gauze.



4 Close the jaws without activating energy.

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

ADDITIONAL IN-SERVICE RECOMMENDATIONS



5 Pull and release the blade lever to clear the blade channel of eschar.



6 Wipe the jaw again before continuing with the surgical procedure.

WARNINGS*:

- Do not activate the device while cleaning the jaws. Do not attempt to deploy the blade while the jaw is open. In both cases, user injury may result.
- Do not use scratch pads or abrasive pads to clean the jaws, or damage to the electrodes may occur, potentially affecting device performance.

* Please refer to page 10 for additional warnings.

Voyant MARYLAND FUSION DEVICE WITH SINGLE-STEP ACTIVATION

ADDITIONAL WARNINGS

Activating Energy

- Do not attempt to seal over clips or staples, as incomplete seals may be formed.
- Do not activate the device when it is in contact with, or in close proximity to, other metal instruments or objects. Burns to the patient and/or surgical team may occur.
- Prior to activating the device, remove any conductive fluid that is in contact with, or in close proximity to, the electrodes of the device. Conductive fluids (e.g., blood, saline, water, etc.) can transfer current, heat, or both, and cause unintended tissue effects.
- Ensure that the jaws of the device are visible prior to and during device activation to avoid unintended tissue effects and/or injury.

Cleaning the Jaws

- Buildup of eschar can negatively affect the device's sealing and cutting performance and/or create embers that present a fire hazard.

Visit appliedmedical.com/voyant/marylandfusion for more information.

The Voyant Fusion devices are bipolar, electrosurgical devices used in procedures where the ligation and division of vessels and tissue bundles is desired.

Please contact your Applied Medical representative for more information on availability. This information is intended for dissemination exclusively to healthcare professionals and is not intended to replace labeling and Instructions for Use (IFU). Please refer to the IFU for the indications, contraindications, warnings, precautions, instructions and other information.

© 2023 Applied Medical Resources Corporation, Rancho Santa Margarita, CA. All rights reserved. Applied Medical, the Applied Medical logo design, and marks designated with the symbol ® are trademarks of Applied Medical Resources Corporation, registered in one or more of the following countries: Australia, Canada, Japan, South Korea, the United States, the United Kingdom, and/or the European Union.

02/2023

472238-EN-B

